


CALIBRATE THERMOMETERS OFTEN


Make sure temperatures are accurate!

ICE WATER METHOD



- Fill a large glass with crushed ice
- Add clean tap water until glass is full
- Stir mixture well
- Put thermometer or probe stem into ice water so the sensing area is completely submerged
- Wait 30 seconds
- Hold the adjusting nut securely with a wrench or pliers and rotate the head of the thermometer until it reads 32°F
- Wait 30 seconds, and recheck temperature

BOILING WATER METHOD



- Bring clean tap water to a boil in a deep pan
- Put thermometer or probe stem into the boiling water so the sensing area is completely submerged
- Wait 30 seconds
- Hold the adjusting nut securely with a wrench or pliers and rotate the head of the thermometer until it reads 212°F (100°C) or the appropriate boiling temperature
- Wait 30 seconds, and recheck temperature

notes...

- Do not let the stem of the thermometer touch the bottom or sides of the glass or pan
- The thermometer stem or probe must remain in the water
- Press the reset button (if available) on a digital thermometer to adjust the readout